

Chlorpyrifos

• Intended Use

For detection of chlorpyrifos in water (groundwater, surface water, well water). For soil, crop, and food use refer to specific application bulletins.

• Principle

The Chlorpyrifos RaPID Assay[®] applies the principles of enzyme linked immunosorbent assay (ELISA) to the determination of chlorpyrifos. The sample to be tested is added, along with an enzyme conjugate, to a disposable test tube, followed by paramagnetic particles with antibodies specific to chlorpyrifos attached. Both the chlorpyrifos (which may be in the sample) and the enzyme labeled chlorpyrifos (the enzyme conjugate) compete for antibody binding sites on the magnetic particles. At the end of an incubation period, a magnetic field is applied to hold the paramagnetic particles (with chlorpyrifos and labeled chlorpyrifos analog bound to the antibodies on the particles, in proportion to their original concentration) in the tube and allow the unbound reagents to be decanted. After decanting, the particles are washed with Washing Solution.

The presence of chlorpyrifos is detected by adding the enzyme substrate (hydrogen peroxide) and the chromogen (3,3',5,5'-tetramethylbenzidine). The enzyme-labeled chlorpyrifos analog bound to the chlorpyrifos antibody catalyzes the conversion of the substrate/ chromogen mixture to a colored product. After an incubation period, the reaction is stopped and stabilized by the addition of acid. Since the labeled chlorpyrifos (conjugate) was in competition with the unlabeled chlorpyrifos (sample) for the antibody sites, **the color developed is inversely proportional to the concentration of chlorpyrifos in the sample**.

• Reagents

1. Chlorpyrifos Antibody Coupled Paramagnetic Particles

The chlorpyrifos antibody (mouse monoclonal anti-chlorpyrifos) is covalently bound to paramagnetic particles, which are suspended in buffered saline with preservative and stabilizers.

30 test kit: one 20 mL vial
100 test kit: one 65 mL vial

2. Chlorpyrifos Enzyme Conjugate

The horseradish peroxidase (HRP) labeled chlorpyrifos analog is diluted in buffered saline with preservative and stabilizers.

30 test kit: one 10 mL vial
100 test kit: one 35 mL vial

3. Chlorpyrifos Standards

Three concentrations (0.22, 1.0, 3.0 ppb) of chlorpyrifos (as chlorpyrifos analog) in buffered saline with preservative and stabilizers are supplied. Each vial contains 2.0 mL.

4. Control

A concentration (approximately 1.8 ppb) of chlorpyrifos (as chlorpyrifos analog) in buffered saline with preservative and stabilizers. A 2.0 mL volume is supplied in one vial.

5. Diluent/Zero Standard

Buffered saline with preservative and stabilizers without any detectable chlorpyrifos.

30 test kit: one 10 mL vial
100 test kit: one 35 mL vial

6. Color Solution

A solution of hydrogen peroxide and 3,3',5,5'-tetramethylbenzidine in an organic base.

30 test kit: one 20 mL vial
100 test kit: one 65 mL vial

7. Stopping Solution

A solution of sulfuric acid (0.5%).

30 test kit: one 20 mL vial
100 test kit: one 60 mL vial

8. Washing Solution

Preserved deionized water.

30 test kit: one 70 mL vial
100 test kit: one 250 mL vial

9. Test Tubes

Polystyrene tubes (36) are packaged in a box.

30 test kit: one 36 tube box
100 test kit: three 36 tube boxes

• Reagent Storage and Stability

Store all reagents at 2-8°C. Do not freeze. Reagents may be used until the expiration date on the box. *The test tubes require no special storage condition and may be stored separately from the reagents to conserve refrigerator space.*

Consult state, local and federal regulations for proper disposal of all reagents.

• Materials Required but Not Provided

In addition to the reagents provided, the following items are essential for the performance of the test:

Pipets*	Precision pipets capable of delivering 250 and 500 uL and a 1.0 mL repeating pipet.
Vortex Mixer*	Thermolyne Maxi Mix, Scientific Industries Vortex Genie, or equivalent

Magnetic Separation Rack*

RPA-I[™] RaPID Analyzer* or equivalent photometer capable of readings at 450 nm

* These items are available from Strategic Diagnostics Inc.

• Sample Information

This procedure is recommended for use with water samples. Other samples may require modifications to the procedure and should be thoroughly validated.

Samples containing gross particulate matter should be centrifuged to remove particles, alternately, filters provided with the RaPID Prep[™] Soil Collection Kit could be used.

Samples which have been preserved with monochloroacetic acid or other acids, should be neutralized with strong base e.g. 6N NaOH, prior to assay.

If the chlorpyrifos concentration of a sample exceeds 3.0 ppb, the sample is subject to repeat testing using a diluted sample. A ten-fold or greater dilution of the sample is recommended with an appropriate amount of Diluent/Zero Standard or Sample Diluent. For example, in a separate test tube make a ten-fold dilution by adding 100 uL of the sample to 900 uL of Diluent/Zero Standard. Mix thoroughly before assaying. Perform the assay according to the Assay Procedure and obtain final results by multiplying the value obtained by the dilution factor e.g. 10.

The presence of the following substances up to 250 ppm were found to have no significant effect on Chlorpyrifos RaPID Assay results: magnesium, calcium, nitrate, thiosulfate, copper, nickel, sulfite, silicates, phosphate and zinc. Sulfate to 10,000 ppm, humic acid to 25 ppm, and mercury, manganese and iron to 100 ppm had no effect in the assay. In addition, sodium chloride concentrations up to 0.5 M showed no effect on results.

• Reagent Preparation

All reagents must be allowed to come to room temperature and the antibody coupled paramagnetic particles should be mixed thoroughly before use.

• Procedural Notes and Precautions

As with all immunoassays, a consistent technique is the key to optimal performance. To obtain the greatest precision, be sure to treat each tube in an identical manner.

Add reagents directly to the bottom of the tube while **avoiding contact between the reagents and the pipet tip**. This will help assure consistent quantities of reagent in the test mixture.

Avoid cross-contaminations and carryover of reagents by using clean pipets for each sample addition and by avoiding contact between reagent droplets on the tubes and pipet tips.

Avoid foam formation during vortexing.

The magnetic separation rack consists of two parts: an upper rack which will securely hold the test tubes and a lower separator which contains the magnets used to attract the antibody coupled paramagnetic particles. During incubations the upper rack is removed from the lower separator so that the paramagnetic particles remain suspended during the incubation. **For separation steps, the rack and the separator are combined to pull the paramagnetic particles to the sides of the tubes.**

To obtain optimum assay precision, it is important to perform the separation steps carefully and consistently. Decant the rack by slowly inverting away from the operator using a smooth turning action so the liquid flows consistently along only one side of the test tube. While still inverted, place the rack on an absorbent pad and allow to drain. Lifting the rack and replacing gently onto the pad several times will ensure complete removal of the liquid from the rim of the tube (technique is demonstrated on training video, available from Strategic Diagnostics Inc.).

Mix the antibody coupled paramagnetic particles just prior to pipetting.

Do not use any reagents beyond their stated shelf life.

Avoid contact of Stopping Solution (sulfuric acid) with skin and mucous membranes. If this reagent comes in contact with skin, wash with water.

• Limitations

The Chlorpyrifos RaPID Assay will detect chlorpyrifos and related compounds to different degrees. Refer to specificity table for data. The Chlorpyrifos RaPID Assay kit provides screening results. As with any analytical technique (GC, HPLC, etc...) positive results requiring some action should be confirmed by an alternative method.

The total time required for pipetting the magnetic particles should be kept to two (2) minutes or less, therefore the total number of tubes that can be assayed in a run should be adjusted accordingly.

• Quality Control

A control solution at approximately 1.8 ppb of chlorpyrifos (as chlorpyrifos analog) is provided with the Chlorpyrifos RaPID Assay kit. It is recommended that it be included in every run and treated in the same manner as unknown samples. Acceptable limits should be established by each laboratory.

• Assay Procedure

Read Reagent Preparation, Procedural Notes and Precautions before proceeding.

1. Label test tubes for standards, control, and samples.

Tube Number	Contents of Tube
1,2	Diluent/Zero Standard, 0 ppb
3,4	Standard 1, 0.22 ppb
5,6	Standard 2, 1.0 ppb
7,8	Standard 3, 3.0 ppb
9	Control
10	Sample 1
11	Sample 2
12	Sample 3

2. Add 250 uL of the appropriate standard, control, or sample.
3. Add 250 uL of Chlorpyrifos Enzyme Conjugate to each tube.
4. Mix the Chlorpyrifos Antibody Coupled Paramagnetic Particles thoroughly and add 500 uL to each tube.
5. Vortex for 1 to 2 seconds minimizing foaming.
6. Incubate for 15 minutes at room temperature.
7. Separate in the Magnetic Separation Rack for **two (2) minutes**.
8. Decant and **gently** blot all tubes briefly in a consistent manner.
9. Add 1 mL of Washing Solution to each tube and allow them to remain in the magnetic separation unit for **two (2) minutes**.
10. Decant and **gently** blot all tubes briefly in a consistent manner.
11. Repeat Steps 9 and 10 an additional time.
12. Remove the rack from the separator and add 500 uL of Color Solution to each tube.
13. Vortex for 1 to 2 seconds minimizing foaming.
14. Incubate for 20 minutes at room temperature.
15. Add 500 uL of Stopping Solution to each tube.
16. Add 1 ml Washing Solution to a clean test tube. Use as blank in Step 17.
17. Read results at 450 nm within 15 minutes after adding the Stopping Solution.

• Results

Manual Calculations

1. Calculate the mean absorbance value for each of the standards.
2. Calculate the %B/Bo for each standard by dividing the mean absorbance value for the standard by the mean absorbance value for the Diluent/Zero Standard.
3. Construct a standard curve by plotting the %B/Bo for each standard on vertical logit (Y) axis versus the corresponding chlorpyrifos concentration on horizontal logarithmic (X) axis on the graph paper provided.
4. %B/Bo for controls and samples will then yield levels in ppb of chlorpyrifos by interpolation using the standard curve.

(Contact SDI for detailed application information on specific photometers.)

RPA-I RaPID Analyzer

Using the RPA-I RaPID Analyzer, calibration curves can be automatically calculated and stored. Refer to the RPA-I operating manual for detailed instructions. To obtain results from the Chlorpyrifos RaPID Assay on the RPA-I the following parameter settings are recommended:

Data Reduct : Lin. Regression
Xformation : Ln/LogitB
Read Mode : Absorbance
Wavelength : 450 nm
Units : PPB
Rgt Blk : 0

Calibrators:

of Cals : 4
of Reps : 2

Concentrations:

#1: 0.00 PPB
#2: 0.22 PPB
#3: 1.00 PPB
#4: 3.00 PPB

Range : 0.10 - 3.00
Correlation : 0.990
Rep. %CV : 10%

• Expected Results

In a study of 151 water samples, the Chlorpyrifos RaPID Assay was shown to correlate well with gas chromatography ($r = 0.985$).

• Performance Data

Precision

The following results were obtained:

Control	1	2	3	4
Replicates	5	5	5	5
Days	5	5	5	5
n	25	25	25	25
Mean (ppb)	0.16	0.54	1.10	1.92
% CV (within assay)	13.0	8.4	5.3	5.1
% CV (between assay)	6.2	11.3	10.0	9.5

Sensitivity

The Chlorpyrifos RaPID Assay has an estimated minimum detectable concentration, based on a 90% B/Bo of 100 ppt.

Recovery

Four (4) samples, including a municipal water source, drinking water from a local well, a sample from a local pond, and a small creek, were spiked with various levels of chlorpyrifos and then assayed using the Chlorpyrifos RaPID Assay. The following results were obtained:

Amount of Chlorpyrifos Added (ppb)	Mean (ppb)	S.D. (ppb)	% Recovery
0.35	0.33	0.02	93
0.70	0.68	0.03	97
1.25	1.31	0.09	105
2.50	2.34	0.16	94
Average			97

Specificity

The cross-reactivity of the Chlorpyrifos RaPID Assay for various analogues can be expressed as the least detectable dose (LDD) which is estimated at 90% B/Bo, or as the dose required to displace 50% (50% B/Bo).

Compound	LDD (ppb)	50% B/Bo (ppb)
Chlorpyrifos	0.10	0.94
Diazinon	0.12	7.56
Chlorpyrifos-methyl	0.14	3.58
Pirimiphos-ethyl	0.32	57.1
Terbufos	10.3	2180
Pirimiphos-methyl	19.8	688
Disulfoton	26.2	688
Fenitrothion	159	4590
3,5,6-Trichloro-2-pyridinol	>10,000	>10,000

The following compounds demonstrated no reactivity in the Chlorpyrifos RaPID Assay at concentrations up to 10,000 ppb: alachlor, aldicarb, atrazine, benomyl, butachlor, butylate, captan, carbaryl, carbendazim, carbofuran, 2,4-D, 1,3-dichloropropene, dinoseb, lindane, metolachlor, metribuzin, pentachlorophenol, picloram, pirimicarb, profenofos, propachlor, thiabendazole, and thiophanate-methyl.

• Assistance

For ordering or technical assistance contact:
Strategic Diagnostics Inc.
111 Pencader Drive
Newark, Delaware 19702-3322 USA
Phone(800)544-8881
Fax(302)456-6782
www.sdix.com
techservice@sdix.com

• Availability

Strategic Diagnostics Inc.
Chlorpyrifos RaPID Assay*
100 Test Kit
Chlorpyrifos Sample Diluent

*U. S. Patent No. 5,576,187

Z00251

R020498